

KARNATAKA ANTIBIOTICS & PHARMACEUTICALS LIMITED

(A Government of India Enterprise)

ENQUIRY REF. No.	KAPL/CG/011/0180
DATE DUE DATE	09.05.2025
	16/05/2025 (13.00HRS)

Dear Sir.

Please submit your lowest and competitive offer in a SEALED ENVELOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF. NO., DATE and DUE DATE on it/ OR MAIL, with other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc., for below mentioned material/s

	The materials	THE DESCRIPTION	UOM	QTY.
SL. NO.	ITEM CODE	ITEM DESCRIPTION	NOS	01
01	QCRFG002	REFRIGERATOR SINGLE DOOR WITH FREEZER 190L	post or	

Please ensure that your offer reaches us on or before Due Date by courier OR Speed post or By hand in sealed cover only to below office address:

M/s. Karnataka Anitibiotics and Pharmaceuticals Limited Plot No.37, Arka The Business Centre ,NTTF Main Road, Peenya Industrial Area 2nd Phase ,Bengaluru-560058 ph. No.080-23571590

OTHER TERMS:

1. F.O.R TERMS

2. GST %

NOTE:

3. PACKING & FORWARDING CHARGES

4. CREDIT PERIOD

5. DELIVERY OFFERED

6.ATTACHED PAGES

: DOOR DELIVERY

: PLEASE SPECIFY : NOT APPLICABLE

: 30 DAYS

: 02 PAGES

1).IF YOU ARE NOT PARTICIPATING IN THE TENDER PLEASE SEND A REGRET LETTER .

2). VENDER HAS TO QUOTE AS PER OUR TENDER IN YOUR COMPANY LETTER HEAD.

3).QUOTATION MUST BE SUBMITTED IN TWO SEALED COVERS (TECHNICAL&COMMERCIAL PRICE BID)SEPARATELY AND IN ONE ENVELOP OR ELSE YOUR PROPOSAL WILL NOT BE CONSIDERED.

IF YOU NEED ANY CLARIFICATION, PLEASE CONTACT US.

Thanking you,

Yours faithfully, For KARNATAKA ANTIBIOTICS & PHARMACEUTICALS LIMITED

YUVARAJA M DEPUTY MANAGER PURCHASE DEPT MOB:9945317873



Karnataka Antibiotics and Pharmaceuticals Limited, Bengaluru

Revised User Requirement Specification

QUALITY CONTROL DEPARTMENT LABORATORY REFRIGERATOR

1. Purpose

The purpose of the equipment is to preserve the working and Reference drug standards under the controlled Temperature.

2. Description and Quantity

Description	Number of Units required
Single door Refrigerator with freezer	1 unit

3. User Specifications

#	Features	Specification	
1,	Fridge temperature range	+2°C to +8°C	
2.	Maximum ambient temperature	+ 40°C	
3.	Freezer temperature range	-5°C to -25°C	
4.	Door	Automatic door closure and solid insulated door with lock & Key.	
5.	Shelves	Adjustable shelves	
6.	Air circulation	Air forced cooling	
7.	Capacity	Around 190 Liters	
8.	Required facilities	 Frost free, Inner LED light. Defrost system by auto or Manual. Not less than 3 star Caster wheels with brake for easy movements 	

4. Warranty

Not less than 3 star, maximum 10 years warranty or as per PO terms and conditions.

5. Documentation

Supplier shall provide the following documents in English language

Operational and maintenance manual

6. Installation and commissioning

The vendor/ supplier shall ensure the correct positioning to the satisfaction of KAPL and be
responsible for the installation of the equipment. The vendor/supplier shall also be responsible for
the re assembly of sections or components of the packaged equipment, if any, including wiring,
spares and any other equipment and components shipped separately or in sections

7. Training

The vendor /supplier shall provide training to the end user at site at no extra cost.
 The qualified representative should trail the end users.