



**KARNATAKA ANTIBIOTICS &
PHARMACEUTICALS LIMITED**

(A Government of India Enterprise)

ENQUIRY REF No:	KAPL/QAD/020/0630
DATE	17-07-2025
DUE DATE	21-07-2025 (13:00HRS)

Dear Sir,

Please submit your lowest and competitive offer in a SEALED ENVELOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF. NO., DATE and DUE DATE on it/ OR MAIL, with other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc., for below mentioned material/s

SL NO.	ITEM CODE	ITEM DESCRIPTION	UOM	QTY
01	QSPHPL130	HPLC VIALS 2 ML WITH SCREW CAP SEPTA	NOS	2000

- 1) Please ensure that your offer reaches us on or before Due Date by courier OR speed post Or You can also mail us to our email: puren@kaplindia.com
- 2) Please send your quotation mentioning item code

OTHER TERMS:

- | | | |
|----|------------------------------|------------------|
| 1. | F.O.R TERMS | : DOOR DELIVERY |
| 2. | GST % | : PLEASE SPECIFY |
| 3. | PACKING & FORWARDING CHARGES | : NOT APPLICABLE |
| 4. | CREDIT PERIOD | : 30 DAYS |
| 5. | DELIVERY OFFERED | : |
| 6. | ATTACHED PAGES | :03 |

NOTE: IN CASE YOU ARE NOT QUOTING PLEASE SEND THE REGRET LETTER.

Thanking you,

Yours faithfully,
For KARNATAKA ANTIBIOTICS
& PHARMACEUTICALS LIMITED

YUVARAJA M
DEPUTY MANAGER PURCHASE DEPT
MOB: 9945317873



**KARNATAKA ANTIBIOTICS AND PHARMACEUTICALS
LIMITED, BENGALURU**

QUALITY CONTROL DEPARTMENT

User Requirement Specifications

Material Description: CLEAR HPLC VIAL WITH PRE SLIT SEPTA

URS Number: QC/URS/001/0725

1. Description and Quantity

Material Description	CLEAR HPLC VIAL WITH PRE SLIT SEPTA
Item Code	FDACC010
Quantity/Box	100 vial/20 Boxes

2. User Specifications

For Clear Glass with writing patch		
1.	Material of Construction	USP Type 1 Borosilicate Glass
2.	Glass Properties	Acid-resistant, Alkali resistant and High temperature resistant
3.	O.D. × H × I.D.	12 mm × 32 mm × 6 mm
4.	Fitting	Thread for 9 mm
5.	Colour	Clear
6.	Style	Screw Top
7.	Write-On Spot and Fill Lines	Yes
8.	Capacity	2 mL
9.	Volume (Metric) Usable	1.5 mL
10.	Tabulation	Neck with tabulation
11.	Closure Type	Screw
12.	Base Style	Flat Bottom
For Screw Cap with Pre Slit Septa		
13.	Material of Construction of Screw Cap	PTFE
14.	Cap Color	Blue
15.	Material of Construction of Pre Slit Septum	Silicone (Pre Slit)
16.	Septum Color	White/red
17.	Septa Layering	Bi-Layer
18.	Diameter	9.00 mm




**KARNATAKA ANTIBIOTICS AND PHARMACEUTICALS
LIMITED, BENGALURU**

FORMULATION DEVELOPMENT

User Requirement Specifications

Material Description: CLEAR HPLC VIAL WITH PRE SLIT SEPTA

URS Number: QC/URS/001/0725

19.	Height	6.2 mm
20.	Image for Illustration	 2mL Autosampler Vials with Screw Cap
21.	Make Preference	1. Waters 2. Agilent 3. Borosil 4. Shimadzu 5. Thermo Scientific 6. PerkinElmer 7. Axiva 8. Rankem