



**KARNATAKA ANTIBIOTICS &
PHARMACEUTICALS LIMITED**

(A Government of India Enterprise)

ENQUIRY REF. No.	KAPL/QAD/020/0963
DATE	01.09.2025 [03/09/25]
DUE DATE	09/09/2025 (13.00HRS)

Dear Sir,

Please submit your lowest and competitive offer in a SEALED ENVELOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF. NO., DATE and DUE DATE on it with other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc., for below mentioned material/s

SL. NO.	ITEM CODE	ITEM DESCRIPTION	UOM	QTY
01	RDACC001	NYLON SYRINGE FILTERS 0.45 MICRON -13 MM (AXIVA)	NOS	05
02	FDACC003	SYRINGE FILTER NYLON 0.45µ DIAMETER 25 MM (AXIVA)	NOS	30

Please ensure that your offer reaches us on or before Due Date by courier OR Speed post or
By hand in sealed cover only to below office address:

M/s. Karnataka Antibiotics and Pharmaceuticals Limited Plot No.37, Arka The Business
Centre ,NTTF Main Road, Peenya Industrial Area 2nd Phase ,Bengaluru-560058 ph. No.080-23571590

OTHER TERMS:

- | | |
|---------------------------------|------------------|
| 1. F.O.R TERMS | : DOOR DELIVERY |
| 2. GST % | : PLEASE SPECIFY |
| 3. PACKING & FORWARDING CHARGES | : NOT APPLICABLE |
| 4. CREDIT PERIOD | : 30 DAYS |
| 5. DELIVERY OFFERED | : |
| 6. ATTACHED PAGES | : 01 |

NOTE:

1).IF YOU ARE NOT PARTICIPATING IN THE TENDER PLEASE SEND A REGRET LETTER .

2).VENDER HAS TO QUOTE AS PER OUR TENDER IN YOUR COMPANY LETTER HEAD.

3).QUOTATION MUST BE SUBMITTED IN TWO SEALED COVERS (TECHNICAL&COMMERCIAL
/PRICE BID)SEPARATELY AND IN ONE ENVELOP OR ELSE YOUR PROPOSAL WILL NOT BE
CONSIDERED.

IF YOU NEED ANY CLARIFICATION , PLEASE CONTACT US.

Thanking you,

Yours faithfully,
For KARNATAKA ANTIBIOTICS
& PHARMACEUTICALS LIMITED


YUVARAJA M
DEPUTY MANAGER PURCHASE DEPT
MOB: 9945317873



KARNATAKA ANTIBIOTICS AND PHARMACEUTICALS LIMITED, BENGALURU

RESEARCH AND DEVELOPMENT (R&D)

User Requirement Specifications

Material Description: Nylon Syringe Filters 0.45 µm-13 mm

URS Number: RD/URS/020/0825/V1

1. Description and Quantity

Material Description	Nylon Syringe Filters 0.45 µm-13 mm
Item Code	RDACC001
Quantity/Box	100 Nos.

2. User Specifications

1.	Make	Axiva Sichem Pvt. Ltd
2.	Catalogue No.	SFNY13XB
3.	Media	NYLON
4.	Filtration Rating	0.45 µm
5.	Diameter	13 mm
6.	Integrity Test (In Water)	Bubble Point (BP) ≥32 psi
7.	Housing Leakage Test (Blocked filter from bottom side, dipped in water and exposed to 70 psi)	No bubble found
8.	Initial water flow rate	26.5 mL/min at 0.7 kg/cm ² at 25°C through pressure vessel
9.	HPLC Extractable Test	No extraneous peaks found



KARNATAKA ANTIBIOTICS AND PHARMACEUTICALS LIMITED, BENGALURU

RESEARCH AND DEVELOPMENT (R&D)

User Requirement Specifications

Material Description: Nylon Syringe Filters 0.45 µm -25 mm

URS Number: RD/URS/019/0825/V1

1. Description and Quantity

Material Description	Nylon Syringe Filters 0.45 µm -25 mm
Item Code	FDACC003
Quantity/Box	100 Nos.

2. User Specifications

1.	Make	Axiva SicheM Pvt. Ltd
2.	Catalogue No.	SFNY25XB
3.	Media	NYLON
4.	Filtration Rating	0.45 µm
5.	Diameter	25 mm
6.	Integrity Test (In Water)	Bubble Point (BP) ≥32 psi
7.	Housing Leakage Test (Blocked filter from bottom side, dipped in water and exposed to 70 psi)	No bubble found
8.	Initial water flow rate	75.5 mL/min at 0.7 kg/cm ² at 25°C through pressure vessel
9.	HPLC Extractable Test	No extraneous peaks found