



**KARNATAKA ANTIBIOTICS &
PHARMACEUTICALS LIMITED**

(A Government of India Enterprise)

ENQUIRY REF. No.	KAPL/CHE/018/1152
DATE	06/10/2025
DUE DATE	08/10/2025 TILL 23.59 HRS

Dear Sir,

Please submit your lowest and competitive offer in an SEALED ENVELOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF NO, DATE and DUE DATE on it/ OR MAIL, with other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc. for the below materials.

SL NO	ITEM CODE	ITEM DESCRIPTION	UOM	QTY
01	FDACC003	Syringe Filter Nylon 0.45µm Diameter 25mm (AXIVA)	NOS	100

Boxes .

NOTE: URS COPY ATTACHED.

OTHER TERMS:

- | | |
|---------------------------------|-----------------------------------|
| 1. F.O.R TERMS | : DOOR DELIVERY PEENYA, BANGALORE |
| 2. GST% | : PLEASE SPECIFY |
| 3. PACKING & FORWARDING CHARGES | : NOT APPLICABLE |
| 4. CREDIT PERIOD | : 45 DAYS REQUESTED, |
| 5. DELIVERY OFFERED | |

INCASE YOU ARE UNABLE TO QUOTE PLEASE SEND REGRET LETTER

For KARNATAKA ANTIBIOTICS
& PHARMACEUTICALS LIMITED

**AUTHORISED SIGNATORY
PURCHASE DEPT.**



**KARNATAKA ANTIBIOTICS AND PHARMACEUTICALS
LIMITED, BENGALURU**

QUALITY CONTROL DEVELOPMENT

User Requirement Specifications

Material Description: Nylon Syringe Filters 0.45 μ m -25 mm

URS Number: QC/URS/025/0925/N1

1. Description and Quantity

Material Description	Nylon Syringe Filters 0.45 μ m -25 mm
Item Code	FDACC003
Quantity/Box	100 Nos.

2. User Specifications

1.	Make	Axiva SicheM Pvt. Ltd
2.	Catalogue No.	SFNY25XB
3.	Media	NYLON
4.	Filtration Rating	0.45 μ m
5.	Diameter	25 mm
6.	Integrity Test (In Water)	Bubble Point (BP) $\geq$ 32 psi
7.	Housing Leakage Test (Blocked filter from bottom side, dipped in water and exposed to 70 psi)	No bubble found
8.	Initial water flow rate	75.5 mL/min at 0.7 kg/cm ² at 25°C through pressure vessel
9.	HPLC Extractable Test	No extraneous peaks found