



**KARNATAKA ANTIBIOTICS &
PHARMACEUTICALS LIMITED**

(A Government of India Enterprise)

ENQUIRY REF. No.	KAPL/QAD/020/2645
DATE	10.02.2026
DUE DATE	16.02.2026 (13.00HRS)

Dear Sir,

Please submit your lowest and competitive offer in a SEALED ENVELOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF. NO., DATE and DUE DATE on it/ with other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc., for below mentioned material/s

SL. NO.	ITEM CODE	ITEM DESCRIPTION	UOM	QTY.
01	FDACC012	NYLON DL SYRINGE FILTER 0.45 MICROMETER – 25 MM	NOS	02

MAKE- AXIVA

Please ensure that your offer reaches us on or before Due Date by courier OR Speed post or By hand in sealed cover only to below office address:

M/s. Karnataka Antibiotics and Pharmaceuticals Limited Plot No.37, Arka The Business Centre ,NTTF Main Road, Peenya Industrial Area 2nd Phase ,Bengaluru-560058 ph. No.080-23571590

OTHER TERMS:

- | | |
|---------------------------------|-----------------------------------|
| 1. F.O.R TERMS | : DOOR DELIVERY TO PEENYA FACTORY |
| 2. GST % | : PLEASE SPECIFY |
| 3. PACKING & FORWARDING CHARGES | : NOT APPLICABLE |
| 4. CREDIT PERIOD | : 30 DAYS |
| 5. DELIVERY OFFERED | : |
| 6. ATTACHED PAGES | : 01 |

NOTE:

1).IF YOU ARE NOT PARTICIPATING IN THE TENDER PLEASE SEND A REGRET LETTER .

2).VENDER HAS TO QUOTE AS PER OUR TENDER IN YOUR COMPANY LETTER HEAD.

3).QUOTATION MUST BE SUBMITTED IN TWO SEALED COVERS (TECHNICAL&COMMERCIAL /PRICE BID)SEPARATELY AND IN ONE ENVELOP OR ELSE YOUR PROPOSAL WILL NOT BE CONSIDERED.

IF YOU NEED ANY CLARIFICATION , PLEASE CONTACT US.

Thanking you,

Yours faithfully,
For KARNATAKA ANTIBIOTICS
& PHARMACEUTICALS LIMITED

YUVARAJA M
DEPUTY MANAGER PURCHASE DEPT



**KARNATAKA ANTIBIOTICS AND PHARMACEUTICALS
LIMITED, BENGALURU**

FORMULATION DEVELOPMENT

User Requirement Specifications

Material Description: Nylon DL Syringe Filters 0.45 µm -25 mm

URS Number: FD/URS/001/0124/V1

1. Description and Quantity

Material Description	Nylon DL Syringe filters 0.45 µm-25 mm
Item Code	FDACC012
Quantity/Box	100 Nos.

2. User Specifications

1.	Make	Axiva Sichem Pvt. Ltd ✓
2.	Catalogue No.	SFGFNY425XB
3.	Media	Nylon DL
4.	Filtration Rating	0.45 µm
5.	Diameter	25 mm
6.	Integrity Test (In water)	Bubble Point (BP) ≥33 psi
7.	Housing Leakage Test (Blocked filter from bottom side, dipped in water and exposed to 70 psi)	No bubble found
8.	Initial water flow rate	76.2 mL/min at 0.7 kg/cm ² at 25°C through pressure vessel
9.	HPLC Extractable Test	No extraneous peaks found