



ENQUIRY REF No:	KAPL/QAD/020/2387
DATE	30.03.2026
DUE DATE	02.04.2026 (13:00HRS)

Dear Sir,

Please submit your lowest and competitive offer in a SEALED ENVELOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF. NO., DATE and DUE DATE on it/ OR MAIL, with other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc., for below mentioned material/s

SL NO.	ITEM CODE	ITEM DESCRIPTION	UOM	QTY
01	FDACC007	HIGH FLOW NYLON MEMBRANE FILTER 0.45µM- 47MM	NOS	05

- 1) Please ensure that your offer reaches us on or before Due Date by courier OR speed post Or you can also mail us to our email: puren@kaplindia.com
- 2) Please send your quotation mentioning item code

OTHER TERMS:

1. F.O.R TERMS : DOOR DELIVERY TO PEENYA FACTORY
2. GST % : PLEASE SPECIFY
3. PACKING & FORWARDING CHARGES : NOT APPLICABLE
4. CREDIT PERIOD : 30 DAYS
5. DELIVERY OFFERED :
6. ATTACHED PAGES : 01

NOTE: IN CASE YOU ARE NOT QUOTING PLEASE SEND THE REGRET LETTER.

Thanking you,

Yours faithfully,

For KARNATAKA ANTIBIOTICS
& PHARMACEUTICALS LIMITED

YUVARAJA M
DEPUTY MANAGER PURCHASE DEPT

MASTER COPY



**KARNATAKA ANTIBIOTICS AND PHARMACEUTICALS
LIMITED, BENGALURU**

FORMULATION DEVELOPMENT

User Requirement Specifications

Material Description: HIGH FLOW NYLON MEMBRANE FILTERS 0.45 µm-47 mm

URS Number: FD/URS/013/0321/V2

1. Description and Quantity

Material Description	High flow nylon membrane filters 0.45 µm-47 mm
Item Code	FDACC007
Quantity/Box	100 Nos.

2. User Specifications

1.	Make	Axiva Sichem Pvt. Ltd
2.	Catalogue No.	170047X/HF
3.	Media	Nylon
4.	Filtration Rating	0.45 µm
5.	Thickness	150 ± 10 µm
6.	Integrity Test	Bubble Point (BP) ≥23 psi
7.	Initial water flow rate	115.5 mL/min at 0.7 kg/cm ² at 25°C through 25 mm S.S filter holder
8.	Sterilization cycles	Maintain integrity after 1 autoclave/steam sterilization cycles at 121°C (15 min; 1 bar)